

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

## ***E-Sampler Real Time Aerosol Monitors ES642 Real Time Aerosol Monitors***

Manufactured by:

### ***MetOne Instruments Inc***

1600 Washington Blvd.  
Grants Pass,  
OR, 97526  
USA

has been assessed by Sira Certification Service  
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Indicative Ambient Particulate Monitors,  
Version 2 dated July 2012**

Certification Range :

PM <sub>2.5</sub>	0-150µg/m <sup>3</sup>
PM <sub>10</sub>	0-150µg/m <sup>3</sup>

Project No. : 16A30898  
Certificate No : Sira MC130241/01  
Initial Certification : 09 December 2013  
This Certificate issued : 15 May 2015  
Renewal Date : 10 December 2018

R Cooper I Eng MInst MC  
Technical Director

MCERTS is operated on behalf of the Environment Agency by

## **Sira Certification Service**

12 Acorn Industrial Park, Crayford Road, Crayford  
Dartford, Kent, UK DA1 4AL  
Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501



*This certificate may only be reproduced in its entirety and without change*  
**Registered Office:** Rake Lane, Eccleston, Chester, UK CH4 9JN

*To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*

## Approved Site Application

*Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net)*

The field test was conducted on a site representative of urban background particulate loading.

## Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

Bureau Veritas Report: AGGX3951580/BV/DH/2764: Test of E-Sampler for use as an Indicative Monitor for PM<sub>10</sub> and PM<sub>2.5</sub> dated July 2012

Sira Report: 16A30898 dated 11/05/2015

## Product Certified

The measuring system consists of the following parts:

- E-Sampler or ES642
- For PM<sub>2.5</sub>, a bespoke PM<sub>2.5</sub> Sharp Cut Cyclone (SCC) designed for a 2 LPM flow rate
- For PM<sub>10</sub>, a bespoke PM<sub>10</sub> Sharp Cut Cyclone (SCC) designed for a 2 LPM flow rate

This certificate applies to all instruments fitted with software version 3693-01 R1.17.1 and serial number J10143 onwards.

Certificate No : Sira MC130241/01  
This Certificate issued : 15 May 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*

## Certified Performance

Test	Result	MCERTS specification
Constancy of the sample volumetric flow	<±3%	Remain constant within ±3% of rated value
Tightness of the sampling system	0.9%	Leakage not to exceed 2% of sampled volume
Maintenance Interval	Monthly	Two weeks

### PM<sub>2.5</sub> E-Sampler 24 hr averages

Test	Result	MCERTS specification
Intra-instrument uncertainty for the candidate method		
All data	0.97 µg/m <sup>3</sup>	≤5µg/m <sup>3</sup>
≥ 12.5 µg/m <sup>3</sup>	1.47 µg/m <sup>3</sup>	
≤ 12.5 µg/m <sup>3</sup>	0.64 µg/m <sup>3</sup>	
Highest resulting uncertainty estimate comparison against data quality objective after slope and intercept correction (Measurement Uncertainty)	23.82%	WCM≤Wd <sub>qo</sub> Measurement uncertainty defined as 50% for indicative instruments

### PM<sub>10</sub> E-Sampler 24 hr averages

Test	Result	MCERTS specification
Intra-instrument uncertainty for the candidate method		
All data	1.55 µg/m <sup>3</sup>	≤5µg/m <sup>3</sup>
≥ 25 µg/m <sup>3</sup>	2.45 µg/m <sup>3</sup>	
≤ 25 µg/m <sup>3</sup>	1.48 µg/m <sup>3</sup>	
Highest resulting uncertainty estimate comparison against data quality objective after slope and intercept correction (Measurement Uncertainty)	20.62%	WCM≤Wd <sub>qo</sub> Measurement uncertainty defined as 50% for indicative instruments

**Note:** The criteria for data split used in the Bureau Veritas report, July 2012, AGGX3951580/BV/DH/2764 were taken from the MCERTS July 2009 specification rather than from the more recent version of MCERTS July 2012. Given that the between sampler/instrument uncertainties calculated for both PM<sub>10</sub> and PM<sub>2.5</sub> are much below the required 5µg/m<sup>3</sup> the relatively minor differences in the processed data split will not result in exceedance of the pass-fail criteria.

Certificate No : Sira MC130241/01  
This Certificate issued : 15 May 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*

## Description

The E-Sampler is a dual technology instrument combining the near-forward light scattering real time particulate measurements with the filter method. Particulate loading of the filter does not affect the light scattering unit performance due to operation of the flow control system.

An internal rotary vane pump draws air the rate of 2 LPM into the sensing chamber of the instrument where it passes through the visible laser light. The instruments are equipped with bespoke 2 LPM cyclones for PM<sub>2.5</sub> and PM<sub>10</sub>.

Ambient temperature and pressure are measured and the actual flow is calculated and controlled by the internal microprocessor making it independent of the filter charge.

Due to its principle of operation it can measure and report data at one to sixtyminute intervals. Longer averaging results are produced by electronic processing of data.

The ES-642 Aerosol Monitor is a compact version of the E-Sampler. The ES-642 and E-Sampler share the same particulate measurement and flow control components. The ES-642 is a real time only measurement so it does not contain a data logger or a filter method.

## General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC130241/01
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC130241/01  
This Certificate issued : 15 May 2015

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*